



Technical requirements:

1. The undeclared tolerance shall be implemented according to GB / T 14486-2008
2. Product appearance requirements: transparent, LDPE
3. Free of bubbles and burrs, outer surface should be smooth, the inside and outside should be clean, no scratch nor damage
4. The liquid in the bottle should not leak after being sealed
5. The whole bottle capacity is 200 ul, and the extrusion amount is 180 ul

GENERAL TOLERANCES	
X.:	±0.3
X.X:	±0.2
X.XX:	±0.1
ANGLE:	±0.5°
UNLESS SPECIFIED	

厦门宝太生物科技有限公司 Xiamen Biotime Biotechnology Co., Ltd.						
型号 MODEL			视图类型 ANGLE		图纸 SIZE	A4
零件名称 PART NAME	single dose buffer		制图 DRAWN BY		日期 DATE	
图号 DWG. NO.			审核 CHECKED BY		日期 DATE	
材质 MATERIAL	PET		确认 APPROVED BY		日期 DATE	
单位 UNIT	mm	比例 SCALE	1:1		日期 DATE	